



Product Service

CERTIFICATE

No. Z1 17 12 22430 052

Holder of Certificate: Alu Rehab ASBedriftsvegen 23
4353 Klepp Stasjon
NORWAY**Production Facility(ies):**

57370

Certification Mark:**Product:****Standard wheelchairs
Manual wheelchair****Model(s):****Netti Mini****Parameters:**

total length incl. footrests:	850 mm
total length without footrests:	720 mm
width:	520 - 620 mm
height:	900 - 1025 mm (push handles)
seat width:	250, 300 and 350 mm
seat depth:	315, 345 and 375 mm

For further Information please see attachment

Tested according to: EN 12183:2014

The product was tested on a voluntary basis and complies with the essential requirements. The certification mark shown above can be affixed on the product. It is not permitted to alter the certification mark in any way. In addition the certification holder must not transfer the certificate to third parties. See also notes overleaf.

Test report no.: 713096379**Valid until:** 2023-01-10**Date,** 2018-01-12
(Torsten Zimmer)

Page 1 of 2