

Alu Rehab AS
Bedriftsvegen 23
4353 Klepp Stasjon
Norway

T: +47 51 78 62 20
post@My-Netti.com

Org.nr: NO 941 181 953 MVA
Bank: 9688 05 00047

TEST DECLARATION

October 2021

Netti III XXHD manual wheelchair:

- is tested and approved according to the European manual wheelchair test standard EN 12183:2014 by accredited test laboratory. Max user weight 200 kg.
Test report is available.
- is approved for being used as a seat in a car. Max user weight is 160 kg.
- is manufactured according to the international quality management standard for medical devices ISO 13485:2016. The quality system is yearly revised by TÜV SÜD. The certificate is available on our home page www.my-netti.com
- is followed by an EU Declaration of Conformity according to MDR 2017/245 also named CE declaration DoC.
The DoC is available on <https://www.my-netti.com>



Sigrun Peerstøe-Kotthaus
Product Safety responsible

Alu Rehab AS