

Alu Rehab AS  
Bedriftsvegen 23  
4353 Klepp Stasjon  
Norway

T: +47 51 78 62 20  
post@My-Netti.com

Org.nr: NO 941 181 953 MVA  
Bank: 9688 05 00047

## TEST DECLARATION

October 2021

### Netti III manual wheelchair:

- is tested and approved according to the European manual wheelchair test standard EN 12183:2014 by accredited test laboratory. Max user weight 145 kg.  
Test report is available.
- is tested according to the international standard crash test standard for wheelchairs ISO 7176-19:2008 and is approved for being used as a seat in a car.  
Max user weight when used as seat in car is 136 kg.
- is manufactured according to the international quality management standard for medical devices ISO 13485:2016. The quality system is yearly revised by TÜV SÜD. The certificate is available on our home page [www.my-netti.com](http://www.my-netti.com)
- is followed by an EU Declaration of Conformity according to MDR 2017/245 also named CE declaration DoC.  
The DoC is available on <https://www.my-netti.com>



Sigrun Peerstøe-Kotthaus  
Product Safety responsible

Alu Rehab AS