


TEST DECLARATION

January 2023

Netti AdaptPro manual wheelchair:

- Is tested and approved according to the European manual wheelchair standard EN 12183:2014 by accredited test laboratory. Max user weight is 135 kg. Test report is available.
- Is tested according to the international standard for crash testing of wheelchairs ISO 7176-19:2008 and is approved for being used as a seat in a car. Max user weight when used as a seat in a car is 135 kg.
- Is manufactured according to the international quality management standard for medical devices ISO 13485:2016. The quality system is yearly revised by TÜV SÜD. The quality system certificate is available on our home page www.my-netti.com.
- Is followed by an EU Declaration of Conformity (CE-declaration) according to MDR (EU) 2017/745. The DoC for each wheelchair model is available on www.my-netti.com



Sigrun Peerstøe-Kotthaus
Netti Product Safety Responsible

Alu Rehab AS