

## TEST DECLARATION

November 2023

### Netti V manual wheelchair:

- Is tested and approved according to the European manual wheelchair standard EN 12183:2022 by accredited test laboratory. Max user weight is 136 kg. Test report is available.
- Is tested according to the international standard for crash testing of wheelchairs ISO 7176-19: 2022 and is approved for being used as a seat in a car. Max user weight when used as a seat in a car is 136 kg.
- Is manufactured according to the international quality management standard for medical devices ISO 13485:2016. The quality system is yearly revised by TÜV SÜD. The quality system certificate is available on our home page [www.my-netti.com](http://www.my-netti.com).
- Is followed by an EU Declaration of Conformity (CE-declaration) according to MDR (EU) 2017/745. The DoC for each wheelchair model is available on [www.my-netti.com](http://www.my-netti.com)



Mike Dongelmans  
Product and Development manager Netti

Alu Rehab AS